

Certificate of Test

Electrical Safety
for
Floor Cable Cover

Date of Issue: 12th July 2013

SGS Reference: ELS 177235/1/C/AB/13

Details of Sample: Model: RFC83H

Date of Receipt: 4th June 2013

Date of Test: 6th June 2013 to 13th June 2013

Test Specification: Clauses 18 & 14 of BS EN 60670-1:2005

(Boxes and enclosures for electrical accessories for household and similar fixed electrical installations – Part 1: General requirements)

Conclusion: The sample tested was found to comply with the Test Specification.



Test Engineer

A Blakett



Counter Signatory

D Carroll

All enquires relating to this certificate must be directed to the Test Engineer

"This document is issued by the Company subject to its General Conditions of Service available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at www.sgs.com/terms_e-document.htm. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law."
"Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 28 days only."